

Fitting guide for AirFit P10 for AirMini




- Align your left (L) and right (R) pillows with your left and right nostrils.



- Pull the headgear straps over your head. Your lower headgear strap should sit comfortably at the back of your head. The top strap should sit comfortably on your crown.



- To achieve a comfortable fit, adjust the headgear straps. For a looser fit, spread the upper and lower straps apart. To tighten, draw them closer together.



- To ensure a comfortable fit, briefly pull your mask away from your nose and reseat the pillows.
- Ensure the nasal pillows fit securely into your nostrils with both pillows positioned upright.



- Your mask is now fitted and ready to use. When the device is running, a light breeze should be felt near the venting area around the tube connection.

Inserting HumidX or HumidX Plus



- To insert your waterless humidifier, simply open your vent module with a twist.



- Insert your choice of humidifier: HumidX or HumidX Plus. It's recommended to begin therapy with HumidX and switch to HumidX Plus if necessary.



- Reassemble your module with another twist.

Note: For complete user instructions, refer to the AirFit P10 for AirMini User Guide.



A simple and easy design

AirFit P10 for AirMini™ offers an expanded feature-set – combining the pocket-sized portability and the new HumidX™ waterless humidification of ResMed's AirMini Solution with the comfort and visual freedom of the existing AirFit P10 mask design.

Available exclusively with AirMini, AirFit P10 for AirMini helps deliver compact, travel-friendly therapy – for patients who want to enjoy the benefits of therapy wherever they go.



